

Certificate of Analysis

Product Name	TE Buffer	
Cat Number	T1120	
Lot Number	2500090004	
Storage	RT	
Test	Specification	Result
Appearance	Colorless clear solution	Colorless clear solution
pH (25℃)	8.0±0.1	8.0
Test Date	02 Sep 2025	
Recommended Retest Period	1 Year	
The results meet the quality requirements		
QC Checker	Niu Hao Ran	QC Manager Quan Jing Yu

